

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 145**

of **CYCLERtest B.V. / Celsiuslabs**
Afdeling Kalibratie "Celsiuslabs"

This annex is valid from: **19-11-2020** to **01-05-2025**

Replaces annex dated: **04-09-2019**

Location(s) where activities are performed under accreditation

Head Office

Rotscherweg 61
 6374 XW
 Landgraaf
 The Netherlands

Location	Abbreviation/ location code
Rotscherweg 61 6374 XW Landgraaf Nederland	LA
On site	OS

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
TE 0 0	Temperature				
TE 4 0	Self Indicating thermometers				
TE 4 1	Temperature sensors with display unit	(30 – 100) °C	0.025 K	In liquid bath	LA
		(30 – 100) °C	0.05 K	In multi-well dry-block calibrator	LA
		(30 - 95) °C	0.2 K	In single-well dry-block calibrator	LA

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2017
Registration number: **K 145**

of **CYCLERtest B.V. / Celsiuslabs**
Afdeling Kalibratie "Celsiuslabs"

This annex is valid from: **19-11-2020** to **01-05-2025**

Replaces annex dated: **04-09-2019**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
TE 13 0	Other temperature enclosures	(30 - 95) °C	0.1 K	- (q)PCR cycler * - Thermocycler * - temperature controlled dry-block incubators *	OS

* Calibrations TE 13.0 on Location Customer are performed by a multi-channel calibration, where the temperature of the thermocycler reaction block is measured simultaneously and dynamically in at least 12.5 % of the wells (or minimum 12 well when reaction block \geq 96 wells) with a measurement interval of maximum 500 ms.

The calibration is performed at least at 3 relevant temperature set points in the range as stated in the scope.

For TE 13.0 the contributions from the device under calibration are excluded from the CMC statement.

The ambient temperature during calibrations in the own laboratory is nominal (23 \pm 5)°C